

Lidocaine in CDCl3 with TMS

Current Data Parameters  
 NAME Oct10-2002  
 EXPNO 9  
 PROCNO 1

F2 - Acquisition Parameters

Date\_ 20021010  
 Time\_ 11.08  
 INSTRUM arx400  
 PROBHD 5 mm QNP 1H  
 PULPROG zg30  
 TD 65536  
 SOLVENT CDCl3  
 NS 8  
 DS 0  
 SWH 8064.516 Hz  
 FIDRES 0.123055 Hz  
 AQ 4.0632820 sec  
 RG 180  
 DW 62.000 usec  
 DE 88.57 usec  
 TE 300.0 K  
 D1 2.0000000 sec  
 P1 7.60 usec  
 SFO1 400.1324008 MHz  
 NUCLEUS 1H

F2 - Processing parameters

SI 65536  
 SF 400.1300187 MHz  
 WDW EM  
 SSB 0  
 LB 0.30 Hz  
 GB 0  
 PC 3.00

1D NMR plot parameters

CX 20.00 cm  
 FIP 10.000 ppm  
 F1 4001.30 Hz  
 F2P -1.000 ppm  
 F2 -400.13 Hz  
 PPMCM 0.55000 ppm/cm  
 HZCM 220.07150 Hz/cm

